



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/509,606	09/27/2004	Henrik Hansson	HW-7319	2293
26294 7590 07/22/2009 TAROLLI, SUNDHEIM, COVELL & TUMMINO L.L.P. 1300 EAST NINTH STREET, SUITE 1700 CLEVELAND, OH 44114				
EXAMINER SCHAPER, MICHAEL T				
ART UNIT		PAPER NUMBER		
3775				
MAIL DATE		DELIVERY MODE		
07/22/2009		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/509,606

Applicant(s)

HANSSON, HENRIK

Examiner

MICHAEL T. SCHAPER

Art Unit

3775

Period for Reply -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on 02 February 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1, 5-14, and 17 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1, 5-14, and 17 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 02 October 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO/S508)
- Paper No(s)/Mail Date 07-15-09
- 4) ☐ Interview Summary (PTO-413)
- Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Response to Arguments

Applicant's arguments filed 2-2-09 have been fully considered but they are not persuasive. As to Applicant arguments that amended claims 1 and 17 are not anticipated by *Purificato* and *Carter*, either alone in combination, Examiner notes that engagement, as opposed to cooperation, of the rotary preventing members is purely manifested at an interface, and no attempt by Applicant to denote this engagement as that of an active and/or direct one (as desired from the Interview Summary of 1-23-09) has been claimed. In addition, FIGS. 5-6 of *Purificato* show rotatability (amendments in claim 1 and 17 are still functional limitations) with respect to the threaded connections detailed therein.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1, 5-10, 12-14 and 17 are rejected under 35 U.S.C. 102(b) as being anticipated by *Purificato* (US Patent 2,631,584).

Purificato discloses a combination of a sleeve (1) having an opening through which a pin [First Interpretation 10 alone; Second Interpretation 10 and 22

together] is located; and an extraction device (**rest of Figure 6**); the extraction device comprising: an inner extraction member/rod [**First Interpretation 18 and 22 together; Second Interpretation 18 alone**]; an outer extraction member [**First Interpretation 13, 14 and 16 together**] ; **Second Interpretation 13 and 16 together**] with a laterally directed handle (**14**) according to the first interpretation; an extraction handle [**First Interpretation 19; Second Interpretation 19 and 14**]; the outer extraction member is manually holdable; the inner extraction member is insertable within the outer extraction member; the outer and inner extraction members are provided with rotary preventing members (inner threads); the rotary preventing members are not circular, but rather helix/spiral or v-shaped; the outer extraction member has an extraction limit (**end of 18**); the inner member has inner threads (**30**) to interact with pin threads (**32**); the inner extraction member has a rear hole (**26**) and a front hole (**32**); the extreme front end of **18** where it meets the knurled knob is tapered inward (relative to another portion, plus the extreme portion considered separate would not be a thread) and the extreme rear is tapered outward (relative to another portion, plus the extreme portion considered separate would not be a thread); and the extraction device consists of only the three members supported by the definition of member from *The American Heritage® Dictionary of the English Language, Fourth Edition* stating a member is: A structural unit, such as a beam or wall, meaning that multiple components/elements/members can be associated with one of the three discussed due to being a structural unit.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 11 is rejected under 35 U.S.C. 103(a) as being unpatentable over Purificato in view of Carter et al (US Patent Application Publication 2001/0056283).

Purificato discloses all of the claimed limitations except for the inner member having lateral holes at the portion with the inner threads. Carter et al teach (**Page 7, Paragraph 0135**) that ports are provided for cleaning the internal portion of a device when an opening already exposed to biological material exists. It would have been obvious to one having ordinary skill in the art at the time the invention was made to further provide for lateral ports to facilitate cleaning and sterilization.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to MICHAEL T. SCHAPER whose telephone number is (571)270-7413. The examiner can normally be reached on M-F, 7:30am-5:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Thomas Barrett can be reached on (571)272-4746. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/M. T. S./
Examiner, Art Unit 3775

/Thomas C. Barrett/
Supervisory Patent Examiner, Art
Unit 3775